

**NEW**

**KYORITSU**  
Quality and reliability is our tradition

# DC MILLIAMP CLAMP LOGGER KEW 2510

**Measures DC 4-20mA signals without breaking the loop**  
**"Advanced model: with recording function!"**



Actual Size


- **Measures 4 to 20 mA instrumentation and process signals without breaking the loop**
- **Memory function stores up to 192000 data useful for monitoring the signals over time and fault finding**
- **Transfer data to PC via Bluetooth**
- **Top class measurement 0.2% accuracy**
- **Ø6mm clamp jaw easy to use in tight places**
- **LED light for illuminating the measurement spot**
- **Analog output terminal to allow the connection to a recorder or a digital multimeter**

## ● KEW 2510 Specifications

DC mA	20/100mA(Auto ranging) ±0.2%rdg±5dgt(0.00mA - 21.49mA) ±1.0%rdg±5dgt(21.0mA - 120.0mA)
Conductor size	φ6mm max.
Analog output	Recorder: DC1000mV against DC100mA
Communication Interface	Bluetooth Ver2.1+EDR Class2
Applicable Standards	IEC 61010-1, Pollution degree 2 IEC 61010-2-032, IEC 61326-1(EMC) IEC 60529 IP40, EN50581(RoHS)
Operating temperature & humidity	-10 - +50°C <85%
Storage temperature & humidity	-20 - +60°C <85%
Power source	R6/LR6(AA) (1.5V) x4 (Alkaline LR6 is recommended.) External supply (AC adapter MODEL8320)
Battery life	Approx. 50 hours continuous with alkaline batteries (with Backlight, LED light and Bluetooth feature OFF)
Dimensions	111(L)x61(W)x46(D)mm : Display unit 104(L)x33(W)x20(D)mm : Sensor 700mm : Sensor Cable
Weight	Approx. 310g(including batteries)
Accessories	8320(AC adapter), KEW Windows for 2510(Software) 9096(Carrying case), LR6(AA)x4, Instruction manual Software installation manual
Optional	7256(Output cord)



### ● Included Accessories




### ● Optional Accessory




### ● Selection Guide

	2500	2510 <b>NEW</b>
DC A 20/100mA	✓	✓
Conductor size φ6mm max	✓	✓
Analog output	✓	✓
Memory		✓
Data transmission (Bluetooth)		✓
Power: Battery	✓	✓
Power: External Supply		✓

### Memory function Communication function



### Analog output terminal for recorder connection



## ! Safety Warnings :

Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

■ For inquires or orders :

## KYORITSU ELECTRICAL INSTRUMENTS WORKS, LTD.

No.5-20,Nakane 2-chome, Meguro-ku, Tokyo, 152-0031 Japan  
 Phone:81-3-3723-0131  
 Fax:81-3-3723-0152  
 E-mail:info@kew-ltd.co.jp  
 Factory: Ehime, Japan

<http://www.kew-ltd.co.jp>